

Pearls

- Always talk to family/ caregivers as they have specific knowledge and skills.
- If using the patient's ventilator bring caregiver knowledgeable in ventilator operation during transport.
- Take patient's ventilator to hospital even if not functioning properly.
- Always use patient's equipment if available and functioning properly.
- Continuous pulse oximetry and End Tidal CO₂ monitoring must be utilized during assessment and transport.
- Unable to correct ventilator problem: Remove patient from ventilator and manually ventilate using BVM.
 - Typical alarms: Low Pressure/ Apnea: Loose or disconnected circuit, leak in circuit or around tracheostomy site. Low Power: Internal battery depleted.
 - High Pressure: Plugged/ obstructed airway or circuit.
 - DOPE: Displaced tracheostomy tube/ ETT, Obstructed tracheostomy tube/ ETT, Pneumothorax and Equipment failure.