

Pearls

- Always talk to family / caregivers as they have specific knowledge and skills.
- If using the patient's ventilator bring caregiver knowledgeable in ventilator operation during transport.
- Always use patient's equipment if available and functioning properly.
- Continuous pulse oximetry and end tidal CO2 monitoring must be utilized during assessment and transport.
- Unable to correct ventilator problem: Remove patient from ventilator and manually ventilate using BVM. Take patient's ventilator to hospital even if not functioning properly.
- Typical alarms: Low Pressure / Apnea: Loose or disconnected circuit, leak in circuit or around tracheostomy site. Low Power: Internal battery depleted.
 - High Pressure: Plugged / obstructed airway or circuit.
 - DOPE: Displaced tracheostomy tube / ETT, Obstructed tracheostomy tube / ETT, Pneumothorax and Equipment failure.

AR 9 This protocol has been altered from the original NCCEP Protocol by the local EMS Medical Director **Airway Respiratory Protocol Section**